



16th November 2020

REFERENCE NUMBER: MGOZ/MPU T 75/2020

Tender for the supply, delivery and installation, testing and commissioning of laboratory equipment for Rural Gozo Directorate

Clarification No. 2

Clarifications

1. The tender submission deadline is being extended and hence the following timetable supersedes the one in the published tender document.

	DATE	TIME*
Last date on which additional information can be issued by the Contracting Authority	18/11/2020	09:30hrs
Deadline for Submission of Tenders (unless otherwise modified in terms of Clause 10.1 of the General Rules Governing Tendering)	26/11/2020	09:30hrs
Tender Opening Session (unless otherwise modified in terms of Clause 10.1 of the General Rules Governing Tendering)	26/11/2020	10:00hrs

* All times Central European Time (CET)

2. Reference is made to Lot 4, Point e., in Section 3 of the tender document which states:

The spectrophotometer must operate in the UV/VIS range of 200 to 900nm, detection range must be at least in the range of 2 - 16,000 ng/μl or better for dsDNA; the photometric range in the 10 mm pathway must be 0.03 - 330 A or better.

Therefore it must possess microvolume reading capabilities.

All other tender documents, conditions and requirements, which are not superseded by this Clarification, remain in place.