



t +356 22156272  
e mpu.mgoz@gov.mt

MINISTRY FOR GOZO  
HEAD OFFICE DG OPERATIONS, ST. FRANCIS  
SQUARE,  
IR-RABAT, GOZO, MALTA

11<sup>th</sup> February 2021

**REFERENCE NUMBER: MGOZ/MPU T 7/2021**

**Simaseed Procurement 7 (2021): Tender for the supply, delivery and installation, testing and commissioning of laboratory equipment at the Government Experimental Farm, Ministry for Gozo, in the context of the Interreg Italia-Malta Simaseed Project.**

**Clarification No. 1**

---

## Clarification

---

**Question 1:** The Microwave Digestor can be provided in different configurations according to

The type of samples  
No of samples per day

Since such information was not provided in the scope of supply and/or technical specifications, may you please provide such information?

**Answer 1:** **In line to the scope of the tender, the type of samples required are soil samples for soil analysis.  
The number of samples per day cannot be predicted. However, the minimum number of samples that the instrument should be able to process at the same time is 10.**

---

All other tender documents, conditions and requirements, which are not superseded by this Clarification, remain in place.