



8th November 2021

REFERENCE NUMBER: SPD3/2021/046

TENDER FOR THE SUPPLY, DELIVERY, INSTALLATION, TESTING AND COMMISSIONING OF A BENCHTOP MICRO-VOLUME SPECTROPHOTOMETER FOR THE RURAL GOZO DIRECTORATE WITHIN THE MINISTRY FOR GOZO

Clarification No. 1

Clarifications

Question 1: Kindly can you advise if stand alone unit with a large LCD (Liquid crystal display) screen non touchscreen is acceptable?

Answer 1: ***As per Clause h in Section 3 of the tender document, 'The spectrophotometer must possess a clear colour LCD display and touchscreen control in English language.'***

All other tender documents, conditions and requirements, which are not superseded by this Clarification, remain in place.

*Sectoral Procurement Unit 3
Sectoral Procurement Directorate*